

DEPARTMENT OF ENVIRONMENTAL QUALITY

M1967 - FCE - 2015 06 02
 AIR QUALITY DIVISION
 FCE Summary Report

Facility : ST. JOHN PROVIDENCE HOSPITAL	SRN : M1967
Location : 16001 WEST NINE MILE RD	District : Southeast Michigan
	County : OAKLAND
City : SOUTHFIELD State: MI Zip Code : 48075	Compliance Status : Compliance
Source Class : SM OPT OUT	Staff : Iranna Konanahalli
FCE Begin Date : 5/26/2014	FCE Completion Date : 6/2/2015
Comments : Hospital Sterilizers are not subject to National Emission Standards for Hazardous Air Pollutants (NESHAP) for Ethylene Oxide Commercial Sterilization and Fumigation Operations (40 CFR, Part 63, Subpart O). Hospital Sterilizers are subject to NESHAP / MA	

List of Partial Compliance Evaluations :

Activity Date	Activity Type	Compliance Status	Comments
05/27/2015	Scheduled Inspection	Compliance	FY 2015 level-2 scheduled inspection of Providence Hospital

Name: Iranna Konanahalli Date: 06/02/2015 Supervisor:

CJE